



**EL Class I**

*Date of issue: 2 July 2026*

*Valid until: 31 December 2027*

*EHEDG hereby declares that the product*

***amperometric oxygen sensors, types InPro6800, InPro6810, InPro6820, InPro6830, InPro6850i, InPro6900, InPro6900i, InPro6950i, InPro6800G, InPro6850iG, InPro6900iG, InPro6950iG and carbon dioxide sensor InPro5000i with PTFE or silicone membrane and silicone or FFKM O-ring***

*from*

*Mettler Toledo GmbH, 15 Im Hackacker, 8902 URDORF, Switzerland*

*has/have been evaluated for compliance and meets/meet the current criteria for Hygienic Equipment Design of closed process applications of the EHEDG*

***Certificate No. EHEDG-C2600046***

Signed \_\_\_\_\_ *Hein Timmerman* \_\_\_\_\_ *President EHEDG*

Signed \_\_\_\_\_ *Karlijn Faber* \_\_\_\_\_ *EHEDG Certification Officer*

**EHEDG**  
**Karspeldreef 8**  
**1101 CJ Amsterdam**  
**Netherlands**

©EHEDG



## APPENDIX 3 EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 28.10.2024

EHEDG File Number: EHEDG-R2400071

Certification Type: EL CLASS I

Applicant: Mettler-Toledo GmbH

Equipment: amperometric oxygen sensors, types InPro6800, InPro6810, InPro6820, InPro6830, InPro6850i, InPro6900, InPro6900i, InPro6950i, InPro6800G, InPro6850iG, InPro6900iG, InPro6950iG and carbon dioxide sensor InPro5000i with PTFE or silicone membrane and silicone or FFKM O-ring

Other essential identification: with process connection InFit761 VARINLINE (Code V01 and V02) and EPDM, FKM, FFKM or Silicone O-ring

### Evaluated by:

Name: Dr. Jürgen Hofmann

Date, Signature: 15.06.2026

### Approved by:

Name: Giampaolo Betta

Date, Signature: 16/06/2026

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for <b>use in Closed Processes</b>
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input checked="" type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
<input type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for <b>EL ASEPTIC</b> Certification.
Criteria	Certification for <b>use in Open Processes</b>
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input type="checkbox"/>	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.

The original of this form will be kept by EHEDG together with the application, the evaluation report, the evidence provided and any other relevant documentation, as listed on the back.

## APPENDIX 3

No.	Description
1.	EHEDG Certificate of Compliance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning-in-place with liquids without dismantling
4.	Appendix 2: conditions for use of the EHEDG Certification Logo
5.	Appendix 3: Equipment evaluation form
6.	Evaluation report of the design of the amperometric oxygen sensors, types InPro6800, InPro6810, InPro6820, InPro6830, InPro6850i, InPro6900, InPro6900i, InPro6950i, InPro6800G, InPro6850iG, InPro6900iG, InPro6950iG and carbon dioxide sensor InPro5000i with PTFE or silicone membrane and silicone or FFKM O-ring, no. 818TUM2026
7.	Drawings of the amperometric oxygen sensors, types InPro6800, InPro6810, InPro6820, InPro6830, InPro6850i, InPro6900, InPro6900i, InPro6950i, InPro6800G, InPro6850iG, InPro6900iG, InPro6950iG and carbon dioxide sensor InPro5000i with PTFE or silicone membrane and silicone or FFKM O-ring, drawing nos: D000344805, 52920947AZ, 52920945AZ, 00079359; original stamped
8.	Test reports of the in-place cleanability test method according to Doc. 2, Test nos. 668/11.02.2019 and 818/03.11.2025.
9.	Cleaning and Installation manual (excerpt) no. 52 403 547 O supplied by the manufacturer
10.	Example of EHEDG Certified Logo Type EL CLASS I