



EL Class I

Date of issue: 21 September 2026

Valid until: 31 December 2027

EHEDG hereby declares that the product


CMFS series flow sensor in sizes 010, 015, 025, 040, 050, 075, 100, 150

from

Emerson Micro Motion Inc, 7070 Winchester Cr. , 80301 Boulder, CO, United States

*has/have been evaluated for compliance and meets/meet the current criteria for
Hygienic Equipment Design of closed process applications of the EHEDG*

Certificate No. EHEDG-C2600054

Signed  President EHEDG
Hein Timmerman

Signed  EHEDG Certification Officer
Karlijn Faber

EHEDG
Karspeldreef 8
1101 CJ Amsterdam
Netherlands

©EHEDG



APPENDIX 3

EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 23.06.2026

EHEDG File Number: EHEDG-R2600021

Certification Type: EL CLASS I

Applicant: Emerson Micromotion, Inc.

Equipment: CMFS series flow sensor in sizes 010, 015, 025, 040, 050, 075, 100, 150

Other essential identification:

Evaluated by:

Name: Mark Morgan, University of Tennessee

Date, Signature: 13.09.2026 *Mark Morgan*

Approved by:

Name: Jurgen Hofmann, AEO

Date, Signature: 17.09.2026 *J. Hofmann*

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for use in Closed Processes
☐	The equipment complies with all applicable HDC in the Guidelines.
☒	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
☐	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for EL ASEPTIC Certification.
Criteria	Certification for use in Open Processes
☐	The equipment complies with all applicable HDC in the Guidelines.
☐	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.



APPENDIX 3

No.	Description
1.	EHEDG Certificate of Compliance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning-in-place with liquids without dismantling
4.	Appendix 2: Conditions for use of the EHEDG Certification Logo
5.	Appendix 3: EHEDG Certification Evaluation Form (this form)
6.	Evaluation Report of the design of the flow sensor, model CMFS-series in sizes 010, 015, 025, 040, 050, 075, 100, 150. UTE 09092026.
7.	Drawings of the CMFS series depicting design details, signed, stamped and dated: Drawing numbers: EB-20014333_AC, EB-20116054_AA.
8.	Test report of the in-place cleanability test method according to Doc. 2 on CMFS100: UTR-100120
9.	Installation and cleaning instructions (excerpt) from the manufacturer Installation Manual 20002158, Rev DQ. January 2024.
10.	Example of EHEDG Certified Logo Type EL CLASS I