

CERTIFICATE OF COMPLIANCE



EL Class I

Date of issue: 22 October 2026

Valid until: 31 December 2027

EHEDG hereby declares that the product


temperature sensors JUMO DTrans T100 and JUMO dTRANS T1000, types 902815 and 902915; electronic temperature switch JUMO DELOS T and JUMO DELOS T02, types 902940 and 902941; hygienic thermowells, type 902812; RTD temperature probes, type 902810

from

JUMO GmbH & Co. KG, Moritz-Juchheim-Str.1, 36039 FULDA, Germany

has/have been evaluated for compliance and meets/meet the current criteria for Hygienic Equipment Design of closed process applications of the EHEDG

Certificate No. EHEDG-C2600064

Signed  President EHEDG
Hein Timmerman

Signed  EHEDG Certification Officer
Karlijn Faber

***EHEDG
Karspeldreef 8
1101 CJ Amsterdam
Netherlands***

©EHEDG



APPENDIX 3

EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 21.03.2024

EHEDG File Number: EHEDG-R2600026

Certification Type: EL CLASS I

Applicant: JUMO GmbH & Co. KG

Equipment: temperature sensors JUMO DTrans T100 and JUMO dTRANS T1000, types 902815 and 902915; electronic temperature switch JUMO DELOS T and JUMO DELOS T02, types 902940 and 902941; hygienic thermowells, type 902812; RTD temperature probes, type 902810

Other essential identification:

Evaluated by:

Name: Dr. Jürgen Hofmann

Date, Signature: 02.10.2026

Approved by:

Name: Irene Llorca

20006744X
MARIA IRENE
LLORCA
(C:G46421988)

Firmado digitalmente por
20006744X MARIA IRENE
LLORCA (C:G46421988)
Fecha: 2026.10.16
18:22:50 +02'00'

Date, Signature: _____

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for use in Closed Processes
☐	The equipment complies with all applicable HDC in the Guidelines.
☒	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
☐	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for EL ASEPTIC Certification.
Criteria	Certification for use in Open Processes
☐	The equipment complies with all applicable HDC in the Guidelines.
☐	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.

The original of this form will be kept by EHEDG together with the application, the evaluation report, the evidence provided and any other relevant documentation, as listed on the back.

APPENDIX 3

No.	Description
1.	EHEDG Certificate of Compliance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning-in-place with liquids without dismantling
4.	Appendix 2: conditions for use of the EHEDG Certification Logo
5.	Appendix 3: Equipment evaluation form
6.	Evaluation report of the design of the temperature sensors JUMO DTrans T100 and JUMO dTRANS T1000, types 902815 and 902915; electronic temperature switch JUMO DELOS T and JUMO DELOS T02, types 902940 and 902941; hygienic thermowells, type 902812; RTD temperature probes, type 902810, no. 789TUM2026
7.	Drawings of the temperature sensors JUMO DTrans T100 and JUMO dTRANS T1000, types 902815 and 902915; electronic temperature switch JUMO DELOS T and JUMO DELOS T02, types 902940 and 902941; hygienic thermowells, type 902812; RTD temperature probes, type 902810, drawing nos. 90281020A44Z001 K000 , 90000015A02Z035 K000, 90281020A44Z002 K000 (page 1+2), 40440000A02Z007 K000 (page 2), 40980300A02Z001 K000; original stamped
8.	Test reports of the in-place cleanability test method according to Doc. 2, Test Nos. 789/29.02.2024 and 820/12.12.2025.
9.	Cleaning and Installation Manual (excerpt) Nos. 90000000T89Z001K000, 00000000T89Z300K000 supplied by the manufacturer
10.	Example of EHEDG Certified Logo Type EL CLASS I