



Date of issue: 27 February 2020

Valid until: 31 December 2024

EL Aseptic Class I

EHEDG hereby declares that the product

Memosens pH-sensor type CPS77D / OCPS77D / CPS47D / OCPS47D / CPS77E-xx6xxxx+* / OCPS77E-xx6xxxxxxxx+* / CPS47E-xx6xxxx+* / OCPS47E-xx6xxxxxxxx+* ISFET

from

Endress+Hauser Conducta GmbH+Co. KG, Dieselstrasse 24, 70839 Gerlingen, Germany

has/have been evaluated for compliance and meets/meet the current criteria for Hygienic Equipment Design of the EHEDG

Certificate No. EHEDG-C2000002

Signed _____

President EHEDG

Hein Timmerman

Signed _____

EHEDG Certification Officer

Karlijn Faber

*EHEDG
Karspeldreef 8
1101 CJ Amsterdam
Netherlands*

©EHEDG

Appendix 3

EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 23.01.2019

EHEDG File Number: EHEDG-C2000002

Certification Type: EL ASEPTIC CLASS I

Applicant: Endress+Hauser Conducta GmbH+Co. KG

Equipment: Memosens pH-sensor type CPS77D / OCPS77D / CPS47D / OCPS47D / CPS77E-xx6xxxx+* / OCPS77E-xx6xxxxxxx+* / CPS47E-xx6xxxx+* / OCPS47E-xx6xxxxxxx+* ISFET

Other essential identification:

Evaluated by:

Name: Dr. Jürgen Hofmann

Approved by:

Name: Dr. Alan Friis

Title: AEO

Date, Signature:

18.02.2020 

| | |
|---|---|
| 1. Results of inspection for compliance with the EHEDG Hygienic Design Criteria. Conclusion: The equipment complies with the criteria. The use of the EHEDG Certification logo is justified: | YES <input type="checkbox"/> MAYBE <input checked="" type="checkbox"/> |
| 2. Evidence for compliance provided and convincing for Certification. Conclusion: The equipment complies with the criteria where possible. The use of the EHEDG Certification logo is justified: | YES <input checked="" type="checkbox"/> |

Signature:



Date: 14.02.2020

The original of this form will be kept by EHEDG together with the application, the inspection report, the evidence provided and any other relevant documentation, as listed on the back.

Appendix 3

| No. | Description |
|-----|--|
| 1. | EHEDG Certificate of Compliance |
| 2. | Contract to use the EHEDG Certification Logo for equipment |
| 3. | Appendix 1: Equipment intended for cleaning-out-of-place with liquids with dismantling and for aseptic applications (proven steam sterilisable and bacteria tight) |
| 4. | Appendix 2: conditions for use of the EHEDG Certification Logo |
| 5. | Appendix 3: Equipment evaluation form |
| 6. | Evaluation report of the design of the Memosens pH-sensor type CPS77D / OCPS77D / CPS47D / OCPS47D / CPS77E-xx6xxxxx+* / OCPS77E-xx6xxxxxxx+* / CPS47E-xx6xxxxx+* / OCPS47E-xx6xxxxxxx+* ISFET, no. 634TUM2020 |
| 7. | Drawing of the Memosens pH-sensor type CPS77D / OCPS77D / CPS47D / OCPS47D / CPS77E-xx6xxxxx+* / OCPS77E-xx6xxxxxxx+* / CPS47E-xx6xxxxx+* / OCPS47E-xx6xxxxxxx+* ISFET, drawing no. 418233-C (3 pages); original stamped |
| 8. | Test report of the in-place cleanability test method, 634.1/19.03.2018 |
| 9. | Test report of the in-place sterilisability test method, 634.2/19.03.2018 |
| 10. | Test report of the bacteria tightness test method, 634.3/19.03.2018 |
| 11. | Example of EHEDG Certified Logo Type EL ASEPTIC CLASS I |