

CERTIFICATE OF COMPLIANCE



EL Class I

Date of issue: 2 December 2021

Valid until: 31 December 2023

EHEDG hereby declares that the product


9600A Series Electromagnetic Flowmeter. Size DN25 - DN150

from

Schneider Electric Systems USA, Inc., 70 Mechanic Street, MA 02035 Foxboro, United States

*has/have been evaluated for compliance and meets/meet the current criteria for
Hygienic Equipment Design of the EHEDG*

Certificate No. EHEDG-C2100073

Signed  _____ *President EHEDG*
Hein Timmerman

Signed  _____ *EHEDG Certification Officer*
Karlijn Faber

*EHEDG
Karspeldreef 8
1101 CJ Amsterdam
Netherlands*

©EHEDG



Appendix 3

EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date : 04.10.2021

EHEDG File Number : EHEDG-C2100014

Certification Type : EL CLASS I

Applicant : Schneider Electric Systems USA, Inc., 70 Mechanic Street, MA 02035 Foxboro, United states.

Equipment : 9600A Series Electromagnetic Flowmeter. Size DN25 - DN150
With EAF 70 EPDM seals

Evaluated by:

Name: A.M. Agterberg, AEO

Approved by:

Name: Alan Friis

Title: AEO

Date, Signature: 24/11-2021 

1. Results of inspection for compliance with the EHEDG Hygienic Design Criteria. Conclusion: The equipment complies with the criteria. The use of the EHEDG Certification logo is justified:	YES <input type="checkbox"/> MAYBE <input checked="" type="checkbox"/>
2. Evidence for compliance provided and convincing for Certification. Conclusion: The equipment complies with the criteria where possible. The use of the EHEDG Certification logo is justified:	YES <input checked="" type="checkbox"/>

Signature:



Date: 22.11.2021

The original of this form will be kept by EHEDG together with the application, the inspection report, the evidence provided and any other relevant documentation, as listed on the back.

Appendix 3

No.	Description:
1.	Type EL-class I Certificate of Compliance with EHEDG HDC
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning in-place with liquids without dismantling
4.	Appendix 2: Conditions for use of the EHEDG Certification Logo
5.	Appendix 3: Equipment Evaluation Form
6.	Declaration of conformity Schneider Electric
7.	EHEDG Hygienic Design Criteria Evaluation Report TUV 89219286.10
8.	Test report of the in-place cleanability test method TUV 89217382.01
9.	Drawings of the Schneider Model 9600A series Electromagnetic Flowmeter (= OPTIFLUX 6000 Electromagnetic Flowmeter), drawing numbers: VRF-0031-I, VRF-0122-I, G.C.H-4123, G.G.G-3775, F.G.U-3776, FZ 4008605801 rev a, FZ 4008605808 rev a, VRF-0244-I, VRF-0252-I.
10.	Extracts of the manual describing installation and cleaning instructions
11	Example of the EHEDG Certification LOGO TYPE EL CLASS I