

CERTIFICATE OF COMPLIANCE



EL Class I

Date of issue: 20 March 2025

Valid until: 31 December 2026

EHEDG hereby declares that the product

***pH-sensors, type InPro325X(i), InPro310X(i) with process connection InFit761
VARINLINE (Code V01 and V02) and EPDM O-ring***

from

Mettler Toledo GmbH, Im Hackacker 15 , 8902 Urdorf, Switzerland

*has/have been evaluated for compliance and meets/meet the current criteria for
Hygienic Equipment Design of closed process applications of the EHEDG*

Certificate No. EHEDG-C2500009

Signed _____ *Hein Timmerman* _____ *President EHEDG*

Signed _____ *Karlijn Faber* _____ *EHEDG Certification Officer*

***EHEDG
Karspeldreef 8
1101 CJ Amsterdam
Netherlands***

©EHEDG



APPENDIX 3 EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 07.06.2024

EHEDG File Number: EHEDG-R2400070

Certification Type: EL CLASS I

Applicant: Mettler-Toledo GmbH

Equipment: pH-sensors, type InPro325X(i), InPro310X(i) with process connection InFit761 VARINLINE (Code V01 and V02) and EPDM O-ring

Other essential identification:

Evaluated by:

Name: Dr. Jürgen Hofmann

Date, Signature: 19.02.2025

Approved by:

Name: Mark Morgan

Date, Signature: 3/18/25 Mark T Morgan

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for use in Closed Processes
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input checked="" type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
<input type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for EL ASEPTIC Certification.
Criteria	Certification for use in Open Processes
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input type="checkbox"/>	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.

APPENDIX 3

No.	Description
1.	EHEDG Certificate of Compliance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning-in-place with liquids without dismantling
4.	Appendix 2: conditions for use of the EHEDG Certification Logo
5.	Appendix 3: Equipment evaluation form
6.	Evaluation report of the design of the pH-sensors, type InPro325X(i), InPro310X(i) with process connection InFit761 VARINLINE (Code V01 and V02) and EPDM O-ring, no. 624TUM2025
7.	Drawings of the pH-sensors, type InPro325X(i), InPro310X(i) with process connection InFit761 VARINLINE (Code V01 and V02) and EPDM O-ring, drawing nos. 00102195, 00046369, 52403340, 52400500; original stamped
8.	Test report of the in-place cleanability test method according to Doc. 2, Test no. 624/21.12.2017.
9.	Cleaning and Installation manual no. 52 403 547 O supplied by the manufacturer
10.	Example of EHEDG Certified Logo Type EL CLASS I