

# CERTIFICATE OF COMPLIANCE



**EL Class I**

**Date of issue: 14 July 2025**

**Valid until: 31 December 2026**

*EHEDG hereby declares that the product*  
***in-line interface and process monitor, type PhaseGuard ST 40, PhaseGuard AD 40***  
***and TurbiGuard AD 40 with sapphire glass and VARINLINE process connection***

*from*

*Sigrist-Photometer AG, Hofurlistrasse 1 , 6373 Ennetbürgen, Switzerland*

*has/have been evaluated for compliance and meets/meet the current criteria for*  
*Hygienic Equipment Design of closed process applications of the EHEDG*

**Certificate No. EHEDG-C2500032**

Signed \_\_\_\_\_ *Hein Timmerman* President EHEDG

Signed \_\_\_\_\_ *Karlijn Faber* EHEDG Certification Officer

**EHEDG**  
**Karspeldreef 8**  
**1101 CJ Amsterdam**  
**Netherlands**

©EHEDG



## APPENDIX 3 EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 30.10.2024

EHEDG File Number: EHEDG-R2500003

Certification Type: EL CLASS I

Applicant: Sigrist-Photometer AG

Equipment: in-line interface and process monitor, type PhaseGuard ST 40, PhaseGuard AD 40 and TurbiGuard AD 40 with sapphire glass

Other essential identification:

**Evaluated by:**

Name: Dr. Jürgen Hofmann

Date, Signature: 12.06.2025

**Approved by:**

Name: Andy Timperley

Date, Signature: 03.07.2025,

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for <b>use in Closed Processes</b>
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input checked="" type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
<input type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for <b>EL ASEPTIC</b> Certification.
Criteria	Certification for <b>use in Open Processes</b>
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input type="checkbox"/>	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.

## APPENDIX 3

No.	Description
1.	EHEDG Certificate of Compliance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning-in-place with liquids without dismantling
4.	Appendix 2: conditions for use of the EHEDG Certification Logo
5.	Appendix 3: Equipment evaluation form
6.	Evaluation report of the design of the in-line interface and process monitor, type PhaseGuard ST 40, PhaseGuard AD 40 and TurbiGuard AD 40 with sapphire glass, no. 802TUM2025
7.	Drawings of the in-line interface and process monitor, type PhaseGuard ST 40, PhaseGuard AD 40 and TurbiGuard AD 40 with sapphire glass, drawing no. PT_EHEDG-001.; original stamped
8.	Test report of the in-place cleanability test method according to Doc. 2, Test no. 802/31.10.2024.
9.	Cleaning and Installation manual no. 16794E   V1   05-2025 supplied by the manufacturer
10.	Example of EHEDG Certified Logo Type EL CLASS I