

CERTIFICATE OF COMPLIANCE



EL Class I

Date of issue: 16 March 2026

Valid until: 31 December 2027

*EHEDG hereby declares that the product
oxygen sensor Memosens, type COS81E in C- and U-shape with silicone diaphragm
and EPDM O-ring*

from

Endress+Hauser Conducta GmbH+Co. KG, Dieselstrasse 24 , 70839 Gerlingen, Germany

*has/have been evaluated for compliance and meets/meet the current criteria for
Hygienic Equipment Design of closed process applications of the EHEDG*

Certificate No. EHEDG-C2600021

Signed _____ *Hein Timmerman* _____ *President EHEDG*

Signed _____ *Karlijn Faber* _____ *EHEDG Certification Officer*

*EHEDG
Karspeldreef 8
1101 CJ Amsterdam
Netherlands*

©EHEDG



APPENDIX 3 EHEDG Certification – Equipment Evaluation Form

Design Evaluation Date: 02.02.2026

EHEDG File Number: EHEDG-R2600017

Certification Type: EL CLASS I

Applicant: Endress+Hauser Conducta GmbH+Co. KG

Equipment: oxygen sensor Memosens, type COS81E in C- and U-shape with silicone diaphragm and EPDM O-ring

Other essential identification:

Evaluated by:

Name: Dr. Jürgen Hofmann

Date, Signature: 26.02.2026

Approved by:

Name: Dr. Nicolas Rossi

Date, Signature: 27/02/26

The use of the EHEDG Certification logo is justified based on the results of the design evaluation, inspection, and testing (as applicable) of the equipment for compliance with the current EHEDG Hygienic Design Criteria (HDC):

Criteria	Certification for use in Closed Processes
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input checked="" type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2.
<input type="checkbox"/>	Evidence for compliance required and provided by in-place cleanability test method according to EHEDG Doc. 2, in-place sterilisability test method according to EHEDG Doc. 5, and bacteria tightness test according to EHEDG Doc. 7 for EL ASEPTIC Certification.
Criteria	Certification for use in Open Processes
<input type="checkbox"/>	The equipment complies with all applicable HDC in the Guidelines.
<input type="checkbox"/>	Evidence for compliance required and provided by OPC cleanability test method according to EHEDG Doc. 57.

APPENDIX 3

No.	Description
1.	EHEDG Certificate of Compliance
2.	Contract to use the EHEDG Certification Logo for equipment
3.	Appendix 1: Equipment intended for cleaning-in-place with liquids without dismantling
4.	Appendix 2: conditions for use of the EHEDG Certification Logo
5.	Appendix 3: Equipment evaluation form
6.	Evaluation report of the design of the oxygen sensor Memosens, type COS81E in C- and U-shape with silicone diaphragm and EPDM O-ring, no. 583TUM2026
7.	Drawings of the oxygen sensor Memosens, type COS81E in C- and U-shape with silicone diaphragm and EPDM O-ring, drawing no. 961007030 (page 1 of 3); original stamped
8.	Test report of the in-place cleanability test method according to Doc. 2, Test nos. 522/25.10.2015 and 583/26.05.2017.
9.	Cleaning and Installation manual (excerpt) nos. BA02066C/07/EN/03.25-00 and SD02751C/07/EN/02.23-00 supplied by the manufacturer
10.	Example of EHEDG Certified Logo Type EL CLASS I